



MISSOURI DEPARTMENT OF HEALTH
STATE PUBLIC HEALTH LABORATORY
DATAMASTER MAINTENANCE REPORT

RECEIVED
Missouri Breath Alcohol Program
By Carol Day at 8:14 am, Mar 04, 2010

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send copy to Department of Health; retain original in department file.

DATAMASTER SN <u>201186</u>	DATE OF INSPECTION <u>02-25-10</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>1007 N. ELM ST. ROLLA, MO. 65401</u>	TIME OF INSPECTION <u>1437 HRS.</u>

CHECKLIST: Place a check (✓) to the left of each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unchecked items must be corrected before using instrument.

☒ DIAGNOSTIC CHECK (PRINTOUT ATTACHED)

☒ COMPUTER

☒ DETECTOR

☒ PROGRAM

☒ FILTERS

☒ HEATERS SAMPLE CHAMBER 49 °C

☒ QUARTZ STANDARD

☒ FLOW DETECTOR

☒ CALIBRATION

☒ PUMP HIGH SPEED

☒ PRINTER

☒ INDICATOR LIGHTS

☒ TIME AND DATE

☒ SIMULATOR TEMPERATURE (34 °C ± 0.2°C) 34°C

☒ CALIBRATION CHECK -

Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) (USE RECIRCULATION PUMP)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

TEST 1 ☒ .099

TEST 2 ☒ .098

TEST 3 ☒ .097

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

☒ NUMBER OF REFUSALS, SINCE LAST MAINTENANCE REPORT, AND NUMBER OF BREATH TESTS IN EACH RANGE AS FOLLOWS: (DO NOT INCLUDE SIMULATOR TESTS)

REFUSALS	<u>3</u> (0-.04)	<u>1</u> (.05-.09)	<u>0</u> (.10-.14)	<u>0</u> (.15-.19)	<u>0</u> (Over .19)	<u>1</u>
----------	------------------	--------------------	--------------------	--------------------	---------------------	----------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary)

INSTRUMENT OPERATES WITHIN SPECIFIED GUIDELINES AS SET FORTH BY MISSOURI D.O.H.

MANUFACTURER: REPCO MARKETING, INC. / CONCENTRATION: .100 / LOT # 09001 / EXP. DATE: 04-22-11

INSPECTING OFFICER

SIGNATURE

Kenneth L. Moberly

PRINT NAME

KENNETH L. MOBERLY

TYPE II PERMIT NUMBER/EXPIRATION DATE

820077 / 02-26-10

TELEPHONE NUMBER

(573) 308-1213

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 09001

EXPIRATION DATE: April 22, 2011 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 09001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1214 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is April 23, 2009.
The expiration date for this lot number is April 22, 2011
at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
ROLLA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201186

02/25/10

ARREST TIME: 14:00

SUBJECT NAME:

MOBERLY/KENNETH/L

DOB: 11/26/63 SEX: M

STATE/D.L.: MO/494825265

ARRESTING OFFICER:

MOBERLY/KENNETH/L

OFFICER I.D.: 998

TESTING OFFICER:

MOBERLY/KENNETH/L

OFFICER I.D.: 998

PERMIT NUMBER: 820077

EXPIRATION DATE: 02/26/10

MISCELLANEOUS DATA:

MONTHLY R.F.I. TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	14:51
INTERNAL STANDARD	VERIFIED	14:51
RADIO INTERFERENCE		

OPERATOR SIGNATURE

Kenneth L. Moberly

CARD STK #
60036

REORDER ALL SUPPLIES FROM N.P.A.S.
2260 NORTH MAIN, MANSFIELD, OH 44903 419-526-6727 (NPAS)

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
ROLLA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201196

02/25/10

TESTING OFFICER:

MOBERLY/KENNETH/L

OFFICER I.D.: 998

PERMIT NUMBER: 020077

EXPIRATION DATE: 02/26/10

MISCELLANEOUS DATA:

MONTHLY MAINTENANCE

--- SUPERVISOR MODE ---

BLANK TEST	.000	14:43
INTERNAL STANDARD	VERIFIED	14:43
EXTERNAL STANDARD	.099	14:44
BLANK TEST	.000	14:44
EXTERNAL STANDARD	.098	14:45
BLANK TEST	.000	14:45
EXTERNAL STANDARD	.097	14:46
BLANK TEST	.000	14:46

N = 3

SIM. = .1

AVG. = .098

OPERATOR SIGNATURE

Kenneth L. Moberly

CARD STK #
60036

REORDER ALL SUPPLIES FROM N.P.A.S.
2260 NORTH MAIN, MANSFIELD, OH 44903 419-526-6727 (NPAS)

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
ROLLA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201186

02/25/10

14:37

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS

SAMPLE CHAMBER: 49c

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz
{|}~

OPERATOR SIGNATURE

Kenneth E. Moberly

CARD STK #
60036

REORDER ALL SUPPLIES FROM N.P.A.S.
2260 NORTH MAIN, MANSFIELD, OH 44903 419-526-6727 (NPAS)

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



KENNETH MOBERLY

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 02/26/08
Number 820077
Expires 02/26/2010

MO 580-0771 (7-88)

Eric C. Olson
Director of State Public Health Laboratory
[Signature]
Director, Department of Health

Lab. 4 (R7-88)

IN THE CIRCUIT COURT OF PHELPS COUNTY, MISSOURI
ASSOCIATE DIVISION

STATE OF MISSOURI, Plaintiff)

vs.)

Case No. _____

_____, Defendant)

AFFIDAVIT PURSUANT TO SECTION 490.692 RSMo

Before me, the undersigned authority, personally appeared KENNETH L. MOBERLY, who, being by me duly sworn, deposes as follows: My name is KENNETH L. MOBERLY. I am of sound mind, capable of making this affidavit, and personally acquainted with the facts herein stated:

I am a "Type II Breathalyzer Operator," authorized by the Missouri Department of Health to carry out maintenance checks on a breathalyzer, and I am the custodian of the records of the breathalyzer for the Rolla Police Department, Rolla, Missouri. Attached hereto are _____ pages of records kept by the Rolla Police Department, Rolla, Missouri, in the regular course of business of the Rolla Police Department, Rolla, Missouri, for an employee or representative of the Rolla Police Department, Rolla, Missouri, with knowledge of the act, event, condition, opinion, or diagnosis recorded to make the records or to transmit information thereof to be included in such record; and the records was made at or near the time of the act, event, condition, opinion, or diagnosis. The records attached hereto are the original or exact duplicates of the original.

Kenneth L. Moberly
Affiant

In witness whereof I have hereunto subscribed my name and affixed my official seal this 2 day of March, 2010.

Jason D. Smith
Notary

My Commission Expires:

5-18-13



JASON D. SMITH
My Commission Expires
May 18, 2013
Phelps County
Commission #09699487